

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102519\
 Data File : VU035486.D
 Acq On : 25 Oct 2019 14:40
 Operator : JC/SP
 Sample : VU1025WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK40

Quant Time: Oct 26 04:17:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 26 04:07:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	210851	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	232578	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	79860	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	134944	4.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.40%
7) Chloroethane-d5	1.93	69	110151	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.60	63	165159	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.70	46	272386	49.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.16%
24) Chloroform-d	5.11	84	186903	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.75	65	105091	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.77	84	402637	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.74	67	130096	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.93	98	332510	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
43) trans-1,3-Dichloropropene-	8.22	79	47614	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.70	63	187762	46.82	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.64%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	100163	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	89180	5.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102519\
Data File : VU035486.D
Acq On : 25 Oct 2019 14:40
Operator : JC/SP
Sample : VU1025WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK40

Quant Time: Oct 26 04:17:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 26 04:07:11 2019
Response via : Initial Calibration

