

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102519\
 Data File : VU035499.D
 Acq On : 25 Oct 2019 20:25
 Operator : JC/SP
 Sample : K5554-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAQ88

Quant Time: Oct 26 05:39:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 26 04:07:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	201098	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	224212	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	78034	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	132196	4.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.00%
7) Chloroethane-d5	1.93	69	110967	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.60	63	164303	3.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.00%
20) 2-Butanone-d5	4.70	46	260028	49.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.24%
24) Chloroform-d	5.11	84	198366	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.75	65	104840	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.77	84	398222	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.74	67	133019	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.93	98	325982	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	8.22	79	46347	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	8.70	63	172008	44.49	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.98%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	99335	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	88329	5.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.00%

Target Compounds					Ovalue
13) Acetone	2.67	43	7520	1.870 ug/L	100
42) Toluene	8.01	91	28388	0.311 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102519\
Data File : VU035499.D
Acq On : 25 Oct 2019 20:25
Operator : JC/SP
Sample : K5554-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAQ88

Quant Time: Oct 26 05:39:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 26 04:07:11 2019
Response via : Initial Calibration

