

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102519\
 Data File : VU035487.D
 Acq On : 25 Oct 2019 15:04
 Operator : JC/SP
 Sample : K5488-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 26 04:23:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 26 04:07:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	208666	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	231072	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	88512	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	140835	5.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.60%
7) Chloroethane-d5	1.93	69	113486	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.60%
11) 1,1-Dichloroethene-d2	2.59	63	169865	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	4.71	46	272531	49.62	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.24%
24) Chloroform-d	5.11	84	201405	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.75	65	111884	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
32) Benzene-d6	5.77	84	413335	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.74	67	136090	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	7.93	98	341266	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
43) trans-1,3-Dichloropropene-	8.22	79	48426	4.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.00%
46) 2-Hexanone-d5	8.68	63	193916	48.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.34%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	103460	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	12.22	152	101052	5.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102519\
Data File : VU035487.D
Acq On : 25 Oct 2019 15:04
Operator : JC/SP
Sample : K5488-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 26 04:23:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 26 04:07:11 2019
Response via : Initial Calibration

