

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
 Data File : VU045379.D
 Acq On : 25 Oct 2021 13:04
 Operator : SY/MD
 Sample : PB140151TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140151TB

Quant Time: Oct 26 04:41:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	220019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	422147	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	392643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	175682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	177647	48.504	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.000%
35) Dibromofluoromethane	5.292	113	127575	47.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	7.903	98	527594	50.141	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	10.645	95	196463	45.225	ug/l	0.01
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.460%
Target Compounds						
					Qvalue	
16) Acetone	2.649	43	19331	8.892	ug/l	96
20) Methylene Chloride	3.035	84	7985	2.439	ug/l	93
43) Isopropyl Acetate	5.925	43	36074	4.105	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
Data File : VU045379.D
Acq On : 25 Oct 2021 13:04
Operator : SY/MD
Sample : PB140151TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140151TB

Quant Time: Oct 26 04:41:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

