

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
 Data File : VU045382.D
 Acq On : 25 Oct 2021 14:16
 Operator : SY/MD
 Sample : M4290-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TH-BD

Quant Time: Oct 26 04:42:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	214358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	407196	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.426	117	375009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	154804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	172155	48.246	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.500%
35) Dibromofluoromethane	5.292	113	125836	48.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	7.902	98	509791	50.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.460%
62) 4-Bromofluorobenzene	10.642	95	178896	42.693	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.380%
Target Compounds						
					Qvalue	
16) Acetone	2.652	43	18384	8.680	ug/l	97
18) Methyl Acetate	2.957	43	9737	1.573	ug/l	97
20) Methylene Chloride	3.038	84	10865	3.406	ug/l	98
43) Isopropyl Acetate	5.922	43	32667	3.854	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
Data File : VU045382.D
Acq On : 25 Oct 2021 14:16
Operator : SY/MD
Sample : M4290-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TH-BD

Quant Time: Oct 26 04:42:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

