

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
 Data File : VU045387.D
 Acq On : 25 Oct 2021 16:14
 Operator : SY/MD
 Sample : M4294-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-GS

Manual Integrations
 APPROVED

MMDadoda
 10/27/2021 9:19:58 AM

Quant Time: Oct 26 04:44:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.356	168	187202	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	6.243	114	348668	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	327730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.825	152	136840	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	144743	46.448	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.900%		
35) Dibromofluoromethane	5.298	113	89965	40.902	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	81.800%		
50) Toluene-d8	7.896	98	443624	51.046	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.100%		
62) 4-Bromofluorobenzene	10.645	95	158186	44.088	ug/l	0.01
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.180%		
Target Compounds						Qvalue
20) Methylene Chloride	3.092	84	10980m	3.941	ug/l	
43) Isopropyl Acetate	5.954	43	37113m	5.113	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
Data File : VU045387.D
Acq On : 25 Oct 2021 16:14
Operator : SY/MD
Sample : M4294-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-GS

Manual Integrations
APPROVED

MMDadoda
10/27/2021 9:19:58 AM

Quant Time: Oct 26 04:44:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

