

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
 Data File : VU045388.D
 Acq On : 25 Oct 2021 16:38
 Operator : SY/MD
 Sample : M4294-11
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-GD

Manual Integrations
 APPROVED

MMDadoda
 10/27/2021 9:20:02 AM

Quant Time: Oct 26 04:44:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.356	168	184980	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	6.247	114	345685	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	324352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	142078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	142432	46.256	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.520%		
35) Dibromofluoromethane	5.295	113	89150	40.881	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	81.760%		
50) Toluene-d8	7.899	98	437011	50.719	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.440%		
62) 4-Bromofluorobenzene	10.642	95	160281	45.057	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.120%		
Target Compounds						Qvalue
20) Methylene Chloride	3.080	84	10507m	3.817	ug/l	
43) Isopropyl Acetate	5.951	43	39277m	5.458	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
Data File : VU045388.D
Acq On : 25 Oct 2021 16:38
Operator : SY/MD
Sample : M4294-11
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-GD

Manual Integrations
APPROVED

MMDadoda
10/27/2021 9:20:02 AM

Quant Time: Oct 26 04:44:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

