

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
 Data File : VU045390.D
 Acq On : 25 Oct 2021 17:52
 Operator : SY/MD
 Sample : M4300-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TH-CS

Quant Time: Oct 26 04:45:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	209893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	398887	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	369774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	162041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	171359	49.045	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.080%
35) Dibromofluoromethane	5.295	113	123183	48.954	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.900%
50) Toluene-d8	7.903	98	502633	50.555	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.100%
62) 4-Bromofluorobenzene	10.636	95	178339	43.447	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.900%
Target Compounds						
					Qvalue	
16) Acetone	2.636	43	11630	5.608	ug/l	98
20) Methylene Chloride	3.047	84	11623	3.721	ug/l	90
43) Isopropyl Acetate	5.915	43	28296	3.408	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102521\
Data File : VU045390.D
Acq On : 25 Oct 2021 17:52
Operator : SY/MD
Sample : M4300-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TH-CS

Quant Time: Oct 26 04:45:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

