

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102618\
 Data File : VU027819.D
 Acq On : 26 Oct 2018 09:06
 Operator : MD/SY
 Sample : VU1026MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK50

Quant Time: Oct 27 01:53:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 27 01:41:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	98341	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	91074	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	41466	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	26207	34.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.50%
7) Chloroethane-d5	1.68	69	21780	36.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.94%
11) 1,1-Dichloroethene-d2	2.28	63	40603	29.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.92%#
21) 2-Butanone-d5	4.18	46	56243	86.25	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.25%
24) Chloroform-d	4.65	84	55257	40.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.34%
26) 1,2-Dichloroethane-d4	5.31	65	42092	44.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.38%
32) Benzene-d6	5.34	84	104366	40.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.48%
36) 1,2-Dichloropropane-d6	6.33	67	38493	41.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.14%
41) Toluene-d8	7.57	98	94474	39.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.98%#
43) trans-1,3-Dichloropropene-	7.85	79	17847	39.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.38%
47) 2-Hexanone-d5	8.31	63	39784	83.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.96%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	49210	41.15	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.30%
64) 1,2-Dichlorobenzene-d4	11.86	152	34462	40.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102618\
Data File : VU027819.D
Acq On : 26 Oct 2018 09:06
Operator : MD/SY
Sample : VU1026MBL01
Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK50

Quant Time: Oct 27 01:53:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 27 01:41:17 2018
Response via : Initial Calibration

