

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102618\
 Data File : VU027831.D
 Acq On : 26 Oct 2018 14:19
 Operator : MD/SY
 Sample : J5668-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETBQ3

Quant Time: Oct 27 02:51:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 27 01:41:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	111997	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	104268	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48602	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	31617	36.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.56%
7) Chloroethane-d5	1.68	69	26434	39.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.80%
11) 1,1-Dichloroethene-d2	2.27	63	46575	29.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.36%#
21) 2-Butanone-d5	4.18	46	74261	100.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.00%
24) Chloroform-d	4.65	84	66969	42.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.50%
26) 1,2-Dichloroethane-d4	5.31	65	48417	44.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.26%
32) Benzene-d6	5.34	84	127211	42.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.68%
36) 1,2-Dichloropropane-d6	6.33	67	47178	44.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.00%
41) Toluene-d8	7.57	98	114419	41.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.54%
43) trans-1,3-Dichloropropene-	7.85	79	21457	41.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.36%
47) 2-Hexanone-d5	8.31	63	53569	98.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.74%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	63411	46.32	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.64%
64) 1,2-Dichlorobenzene-d4	11.86	152	45099	45.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	69058	68.281 ug/L	93
12) 1,1-Dichloroethene	2.28	96	672	1.077 ug/L #	1
13) Acetone	2.33	43	4665	9.018 ug/L	99
20) cis-1,2-Dichloroethene	4.23	96	153173	187.556 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102618\
Data File : VU027831.D
Acq On : 26 Oct 2018 14:19
Operator : MD/SY
Sample : J5668-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETBQ3

Quant Time: Oct 27 02:51:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 27 01:41:17 2018
Response via : Initial Calibration

