

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102619\
 Data File : VU035522.D
 Acq On : 26 Oct 2019 16:44
 Operator : JC/SP
 Sample : K5536-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEK32

Quant Time: Oct 30 02:36:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 30 02:26:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	668549	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	629096	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	227540	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	227760	43.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.38%
7) Chloroethane-d5	1.93	69	197402	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.16%
11) 1,1-Dichloroethene-d2	2.59	63	292498	33.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.40%
21) 2-Butanone-d5	4.69	46	352174	101.24	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.24%
24) Chloroform-d	5.11	84	384782	44.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.16%
26) 1,2-Dichloroethane-d4	5.75	65	250456	45.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.76%
32) Benzene-d6	5.77	84	803955	47.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.70%
36) 1,2-Dichloropropane-d6	6.74	67	265752	48.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.92%
41) Toluene-d8	7.93	98	724498	47.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.30%
43) trans-1,3-Dichloropropene-	8.21	79	125684	48.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.88%
47) 2-Hexanone-d5	8.68	63	241454	119.17	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	119.17%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	353897	45.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.70%
64) 1,2-Dichlorobenzene-d4	12.22	152	219450	49.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.04%

Target Compounds					Ovalue
13) Acetone	2.67	43	16369	4.696 ug/L	99
15) Methyl Acetate	3.00	43	5412	0.945 ug/L	94
16) Methylene chloride	3.09	84	19691	3.740 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102619\
Data File : VU035522.D
Acq On : 26 Oct 2019 16:44
Operator : JC/SP
Sample : K5536-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEK32

Quant Time: Oct 30 02:36:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102619WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 30 02:26:55 2019
Response via : Initial Calibration

