

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102619\
 Data File : VU035517.D
 Acq On : 26 Oct 2019 14:48
 Operator : JC/SP
 Sample : VU1026WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK51

Quant Time: Oct 30 01:34:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 30 02:26:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	628805	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	610552	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	172100	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	237841	47.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.92%
7) Chloroethane-d5	1.93	69	199693	48.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.96%
11) 1,1-Dichloroethene-d2	2.60	63	303086	36.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.16%
21) 2-Butanone-d5	4.69	46	323481	98.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.87%
24) Chloroform-d	5.11	84	387158	47.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.32%
26) 1,2-Dichloroethane-d4	5.75	65	253416	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.64%
32) Benzene-d6	5.77	84	816858	50.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.18%
36) 1,2-Dichloropropane-d6	6.74	67	272845	51.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.52%
41) Toluene-d8	7.93	98	721761	48.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	8.21	79	119012	47.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.50%
47) 2-Hexanone-d5	8.68	63	185292	94.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.23%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	338413	45.17	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.34%
64) 1,2-Dichlorobenzene-d4	12.22	152	178128	53.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102619\
Data File : VU035517.D
Acq On : 26 Oct 2019 14:48
Operator : JC/SP
Sample : VU1026WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK51

Quant Time: Oct 30 01:34:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102619WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 30 02:26:55 2019
Response via : Initial Calibration

