




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102620\
Data File : VU040948.D
Acq On : 26 Oct 2020 16:47
Operator : SY/MD
Sample : L4492-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Quant Time: Oct 27 03:16:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 27 03:00:07 2020
Response via : Initial Calibration



TIC: VU040948.D

(13) Acetone (T)

2.684min (-0.007) 3.30ug/L

response 3567

Ion	Exp%	Act%
43.05	100	100
58.05	12.70	23.24
0.00	0.00	0.00
0.00	0.00	0.00