




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102620\
 Data File : VU040956.D
 Acq On : 26 Oct 2020 19:53
 Operator : SY/MD
 Sample : L4492-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Quant Time: Oct 27 03:17:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 27 03:00:07 2020
 Response via : Initial Calibration



TIC: VU040956.D

(13) Acetone (T)
 2.681min (-0.010) 2.65ug/L
 response 2604

Ion	Exp%	Act%
43.05	100	100
58.05	12.70	35.83#
0.00	0.00	0.00
0.00	0.00	0.00