

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102620\
 Data File : VU040942.D
 Acq On : 26 Oct 2020 14:28
 Operator : SY/MD
 Sample : L4492-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AN1

Quant Time: Oct 27 03:58:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 27 03:00:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	58223	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	60240	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	24859	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	13484	2.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.20%
7) Chloroethane-d5	1.92	69	14383	3.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.60%
11) 1,1-Dichloroethene-d2	2.58	63	24323	2.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.80%#
20) 2-Butanone-d5	4.66	46	86033	53.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.82%
24) Chloroform-d	5.08	84	37544	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
26) 1,2-Dichloroethane-d4	5.72	65	23468	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
32) Benzene-d6	5.74	84	63954	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
36) 1,2-Dichloropropane-d6	6.70	67	23326	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	7.91	98	48998	3.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.20%#
43) trans-1,3-Dichloropropene-	8.19	79	8874	3.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.00%
46) 2-Hexanone-d5	8.64	63	59535	48.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.26%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	22714	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	21774	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds					Ovalue
13) Acetone	2.67	43	10181	9.245 ug/L #	66
16) Methylene chloride	3.06	84	2465	0.514 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102620\
Data File : VU040942.D
Acq On : 26 Oct 2020 14:28
Operator : SY/MD
Sample : L4492-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AN1

Quant Time: Oct 27 03:58:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 27 03:00:07 2020
Response via : Initial Calibration

