

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U102620\
 Data File : VU040972.D
 Acq On : 27 Oct 2020 03:26
 Operator : SY/MD
 Sample : L4492-20
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP8

Manual Integrations
 APPROVED

MMDadoda
 10/27/2020 3:12:57 PM

Quant Time: Oct 27 06:28:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 27 05:24:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	56676	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	58927	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	23495	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	23793	5.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.00%
7) Chloroethane-d5	1.92	69	18856	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.00%
11) 1,1-Dichloroethene-d2	2.58	63	36285	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	4.67	46	82397	52.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.10%
24) Chloroform-d	5.08	84	41653	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.72	65	25902	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.74	84	76646	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.70	67	25882	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.91	98	59882	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	8.19	79	10251	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.64	63	58077	48.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.00%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	23203	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	23667	5.74	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.80%

Target Compounds

					Ovalue
13) Acetone	2.68	43	10987m	10.249	ug/L
14) Carbon disulfide	2.81	76	50139	3.748	ug/L
16) Methylene chloride	3.06	84	2049	0.439	ug/L
47) Tetrachloroethene	8.56	164	2331	0.707	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102620\
Data File : VU040972.D
Acq On : 27 Oct 2020 03:26
Operator : SY/MD
Sample : L4492-20
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP8

Manual Integrations
APPROVED

MMDadoda
10/27/2020 3:12:57 PM

Quant Time: Oct 27 06:28:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 27 05:24:32 2020
Response via : Initial Calibration

