

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102620\
 Data File : VU040953.D
 Acq On : 26 Oct 2020 18:43
 Operator : SY/MD
 Sample : L4492-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AN6

Quant Time: Oct 27 05:16:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 27 03:00:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	53860	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	56121	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	22483	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	13856	3.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.20%
7) Chloroethane-d5	1.92	69	14250	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.00%
11) 1,1-Dichloroethene-d2	2.58	63	23831	2.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.80%#
20) 2-Butanone-d5	4.66	46	83974	56.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.72%
24) Chloroform-d	5.08	84	36706	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.72	65	23571	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
32) Benzene-d6	5.74	84	63354	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
36) 1,2-Dichloropropane-d6	6.70	67	23098	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.91	98	46893	3.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.00%#
43) trans-1,3-Dichloropropene-	8.19	79	8422	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	8.64	63	58350	51.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.32%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	22890	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	20280	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102620\
Data File : VU040953.D
Acq On : 26 Oct 2020 18:43
Operator : SY/MD
Sample : L4492-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AN6

Quant Time: Oct 27 05:16:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 27 03:00:07 2020
Response via : Initial Calibration

