

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
 Data File : VU045400.D
 Acq On : 26 Oct 2021 13:25
 Operator : SY/MD
 Sample : VU1026WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1026WBL01

Quant Time: Oct 26 16:04:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	201820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	380690	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	356818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	157267	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	164640	49.007	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.020%
35) Dibromofluoromethane	5.295	113	120951	50.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.720%
50) Toluene-d8	7.902	98	476728	50.241	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.480%
62) 4-Bromofluorobenzene	10.635	95	172294	43.981	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
Data File : VU045400.D
Acq On : 26 Oct 2021 13:25
Operator : SY/MD
Sample : VU1026WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1026WBL01

Quant Time: Oct 26 16:04:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

