

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
 Data File : VU045408.D
 Acq On : 26 Oct 2021 18:09
 Operator : SY/MD
 Sample : M4360-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-198-200

Quant Time: Oct 27 04:41:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	208264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	392736	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	359910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	153336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	169075	48.770	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.540%
35) Dibromofluoromethane	5.292	113	124872	50.402	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	7.899	98	491556	50.215	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.420%
62) 4-Bromofluorobenzene	10.636	95	172829	42.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.520%
Target Compounds						
					Qvalue	
3) Chloromethane	1.520	50	10098	3.023	ug/l	99
16) Acetone	2.636	43	6967	3.386	ug/l	98
18) Methyl Acetate	2.957	43	7049	1.172	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
Data File : VU045408.D
Acq On : 26 Oct 2021 18:09
Operator : SY/MD
Sample : M4360-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-198-200

Quant Time: Oct 27 04:41:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

