

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
 Data File : VU045411.D
 Acq On : 26 Oct 2021 19:20
 Operator : SY/MD
 Sample : M4360-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-EB-20211022

Quant Time: Oct 27 04:41:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	206502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	394031	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	364679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	160780	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	170088	49.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.960%
35) Dibromofluoromethane	5.292	113	125019	50.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.600%
50) Toluene-d8	7.899	98	494881	50.388	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.780%
62) 4-Bromofluorobenzene	10.635	95	176992	43.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.300%
Target Compounds						Qvalue
3) Chloromethane	1.520	50	8384	2.531	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
Data File : VU045411.D
Acq On : 26 Oct 2021 19:20
Operator : SY/MD
Sample : M4360-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-EB-20211022

Quant Time: Oct 27 04:41:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

