

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
 Data File : VU045412.D
 Acq On : 26 Oct 2021 19:43
 Operator : SY/MD
 Sample : M4360-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-DUP-20211025

Quant Time: Oct 27 04:41:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	213266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	407001	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	374854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	165095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	174184	49.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.120%
35) Dibromofluoromethane	5.292	113	127400	49.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.240%
50) Toluene-d8	7.899	98	514097	50.677	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.360%
62) 4-Bromofluorobenzene	10.636	95	181945	43.442	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.880%
Target Compounds						
					Qvalue	
3) Chloromethane	1.520	50	9284	2.714	ug/l	96
16) Acetone	2.636	43	3718	1.764	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102621\
Data File : VU045412.D
Acq On : 26 Oct 2021 19:43
Operator : SY/MD
Sample : M4360-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-DUP-20211025

Quant Time: Oct 27 04:41:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

