

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102720\
 Data File : VU040985.D
 Acq On : 27 Oct 2020 15:53
 Operator : SY/MD
 Sample : L4493-01DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP9DL

Quant Time: Oct 28 03:16:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 28 02:17:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	55929	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	60394	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	23256	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	23814	5.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.40%
7) Chloroethane-d5	1.92	69	19365	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.58	63	35708	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	4.66	46	81564	52.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.42%
24) Chloroform-d	5.08	84	43463	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.72	65	26607	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
32) Benzene-d6	5.74	84	78491	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.70	67	27214	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.91	98	61550	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	8.19	79	10497	4.60	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.00%
46) 2-Hexanone-d5	8.64	63	58829	47.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.86%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	23033	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	23479	5.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.00%

Target Compounds					Ovalue
47) Tetrachloroethene	8.56	164	20008	5.925 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102720\
Data File : VU040985.D
Acq On : 27 Oct 2020 15:53
Operator : SY/MD
Sample : L4493-01DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP9DL

Quant Time: Oct 28 03:16:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 28 02:17:39 2020
Response via : Initial Calibration

