

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102723\
 Data File : VU055930.D
 Acq On : 27 Oct 2023 17:55
 Operator : MD/SY
 Sample : 05009-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A49K1

Quant Time: Oct 27 23:56:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 25 05:19:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	368878	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	359178	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	173223	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	69444	2.380	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	47.600%	
7) Chloroethane-d5	1.914	69	66519	2.925	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	58.400%#	
11) 1,1-Dichloroethene-d2	2.567	65	32142	2.595	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	51.800%#	
20) 2-Butanone-d5	4.625	46	254719	48.931	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.860%	
24) Chloroform-d	5.059	84	191791	3.547	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	71.000%	
26) 1,2-Dichloroethane-d4	5.698	65	93431	3.547	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.000%	
32) Benzene-d6	5.724	84	342041	3.197	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	64.000%#	
36) 1,2-Dichloropropane-d6	6.689	67	119051	3.754	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	75.000%	
41) Toluene-d8	7.894	98	281302	2.898	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	58.000%#	
43) trans-1,3-Dichloroprop...	8.181	79	39277	3.166	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	63.400%	
46) 2-Hexanone-d5	8.631	63	216317	68.695	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	137.400%#	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	102773	4.450	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	123894	3.834	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	76.600%#	
Target Compounds						
						Qvalue
3) Chloromethane	1.519	50	9451	0.471	ug/L	97
30) Cyclohexane	5.387	56	5967	0.216	ug/L	97
33) Benzene	5.772	78	38056	0.480	ug/L	100
42) Toluene	7.968	91	288160	3.479	ug/L	99
52) Ethylbenzene	9.570	91	70451	0.750	ug/L	99
53) m,p-Xylene	9.692	106	85713	2.443	ug/L	95
54) o-Xylene	10.100	106	34216	0.996	ug/L	94
62) 1,3,5-Trimethylbenzene	11.087	105	16431	0.233	ug/L	92
63) 1,2,4-Trimethylbenzene	11.467	105	63751	0.973	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102723\
Data File : VU055930.D
Acq On : 27 Oct 2023 17:55
Operator : MD/SY
Sample : 05009-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A49K1

Quant Time: Oct 27 23:56:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 25 05:19:13 2023
Response via : Initial Calibration

