

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102820\
 Data File : VU041003.D
 Acq On : 28 Oct 2020 19:40
 Operator : SY/MD
 Sample : VU1028WBL04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK104

Quant Time: Oct 29 02:16:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 29 01:58:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	116289	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	118581	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.81	152	48815	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	39083	4.99	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.80%
7) Chloroethane-d5	1.92	69	31850	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.58	65	18955	5.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.20%
20) 2-Butanone-d5	4.67	46	169699	52.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.10%
24) Chloroform-d	5.08	84	79664	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.72	65	50291	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.75	84	142375	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.70	67	50744	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.91	98	114387	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	8.19	79	19741	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.64	63	116479	50.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.02%
56) 1,1,2,2-Tetrachloroethane-	10.75	84	44326	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
66) 1,2-Dichlorobenzene-d4	12.19	152	44535	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102820\
Data File : VU041003.D
Acq On : 28 Oct 2020 19:40
Operator : SY/MD
Sample : VU1028WBL04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK104

Quant Time: Oct 29 02:16:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 29 01:58:41 2020
Response via : Initial Calibration

