

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102820\
 Data File : VU040996.D
 Acq On : 28 Oct 2020 14:07
 Operator : SY/MD
 Sample : VIBLK01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK01

Quant Time: Oct 29 02:19:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 29 01:58:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	63737	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	65005	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.81	152	26116	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	22658	5.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.60%
7) Chloroethane-d5	1.92	69	19291	5.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.00%
11) 1,1-Dichloroethene-d2	2.58	65	11004	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.20%
20) 2-Butanone-d5	4.66	46	86622	48.48	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.96%
24) Chloroform-d	5.08	84	43031	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.72	65	27810	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.74	84	78906	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.70	67	27666	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.91	98	62733	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	8.19	79	10909	4.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.80%
46) 2-Hexanone-d5	8.64	63	61416	48.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.18%
56) 1,1,2,2-Tetrachloroethane-	10.75	84	23342	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.00%
66) 1,2-Dichlorobenzene-d4	12.19	152	24166	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102820\
Data File : VU040996.D
Acq On : 28 Oct 2020 14:07
Operator : SY/MD
Sample : VIBLK01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK01

Quant Time: Oct 29 02:19:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 29 01:58:41 2020
Response via : Initial Calibration

