

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102821\
 Data File : VU045438.D
 Acq On : 28 Oct 2021 14:36
 Operator : SY/MD
 Sample : VU1028MBL06
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK034

Quant Time: Oct 29 03:11:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 29 03:03:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	319486	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	296506	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	135320	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	125575	48.292	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	96.580%		
7) Chloroethane-d5	1.916	69	107027	51.360	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	102.720%		
11) 1,1-Dichloroethene-d2	2.571	63	167214	36.208	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	72.420%		
21) 2-Butanone-d5	4.623	46	220940	101.529	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	101.530%		
24) Chloroform-d	5.066	84	221158	47.407	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.820%		
26) 1,2-Dichloroethane-d4	5.706	65	152898	49.563	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.120%		
32) Benzene-d6	5.732	84	462177	50.439	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.880%		
36) 1,2-Dichloropropane-d6	6.693	67	149946	50.553	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	101.100%		
41) Toluene-d8	7.899	98	397862	49.773	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	99.540%		
43) trans-1,3-Dichloroprop...	8.182	79	67877	48.707	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.420%		
47) 2-Hexanone-d5	8.632	63	156204	98.660	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	98.660%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	214029	48.084	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	96.160%		
66) 1,2-Dichlorobenzene-d4	12.195	152	142079	51.897	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	103.800%		
Target Compounds					Qvalue	
42) Toluene	7.970	91	6316	0.590	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102821\
Data File : VU045438.D
Acq On : 28 Oct 2021 14:36
Operator : SY/MD
Sample : VU1028MBL06
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK034

Quant Time: Oct 29 03:11:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 29 03:03:47 2021
Response via : Initial Calibration

