

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102823\
 Data File : VU055978.D
 Acq On : 28 Oct 2023 16:53
 Operator : MD/SY
 Sample : 05114-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B62

Quant Time: Oct 30 04:55:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 30 04:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	337974	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	333891	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	155092	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	108947	4.075	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	81.400%	
7) Chloroethane-d5	1.914	69	95377	4.577	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.600%	
11) 1,1-Dichloroethene-d2	2.567	65	50511	4.450	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	89.000%	
20) 2-Butanone-d5	4.628	46	234100	49.082	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	98.160%	
24) Chloroform-d	5.058	84	262641	5.301	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.000%	
26) 1,2-Dichloroethane-d4	5.701	65	112268	4.651	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
32) Benzene-d6	5.724	84	487039	4.897	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
36) 1,2-Dichloropropane-d6	6.689	67	149803	5.081	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.600%	
41) Toluene-d8	7.894	98	417232	4.624	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.400%	
43) trans-1,3-Dichloroprop...	8.177	79	51355	4.453	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.000%	
46) 2-Hexanone-d5	8.631	63	186359	63.664	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	127.320%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	105935	4.934	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.600%	
66) 1,2-Dichlorobenzene-d4	12.190	152	150062	5.186	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.800%	
Target Compounds						Qvalue
25) Chloroform	5.084	83	630192	17.822	ug/L	98
38) Bromodichloromethane	7.103	83	108560	4.420	ug/L	97
49) Dibromochloromethane	8.807	129	17399	1.152	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102823\
Data File : VU055978.D
Acq On : 28 Oct 2023 16:53
Operator : MD/SY
Sample : 05114-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B62

Quant Time: Oct 30 04:55:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 30 04:51:15 2023
Response via : Initial Calibration

