

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102823\
 Data File : VU055974.D
 Acq On : 28 Oct 2023 15:16
 Operator : MD/SY
 Sample : 05114-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B58

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/30/2023
 Supervised By :Mahesh Dadoda 11/01/2023

Quant Time: Oct 30 04:54:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 30 04:51:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	345358	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	344287	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	160856	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	119817	4.386	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	87.800%		
7) Chloroethane-d5	1.914	69	106920	5.021	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	100.400%		
11) 1,1-Dichloroethene-d2	2.567	65	54829	4.727	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	94.600%		
20) 2-Butanone-d5	4.624	46	265736	54.524	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	109.040%		
24) Chloroform-d	5.062	84	286634	5.662	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	113.200%		
26) 1,2-Dichloroethane-d4	5.698	65	123853	5.022	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	100.400%		
32) Benzene-d6	5.724	84	529616	5.164	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	103.200%		
36) 1,2-Dichloropropane-d6	6.689	67	162582	5.348	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	107.000%		
41) Toluene-d8	7.894	98	462958	4.976	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	99.600%		
43) trans-1,3-Dichloroprop...	8.177	79	56173	4.724	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	94.400%		
46) 2-Hexanone-d5	8.631	63	211614	70.108	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	140.220%#		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	116918	5.281	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	105.600%		
66) 1,2-Dichlorobenzene-d4	12.190	152	166174	5.537	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	110.800%		
Target Compounds					Qvalue	
13) Acetone	2.641	43	14048	5.058	ug/L	98
21) 2-Butanone	4.721	43	11927m	2.921	ug/L	
25) Chloroform	5.084	83	663566	18.365	ug/L	99
38) Bromodichloromethane	7.103	83	113786	4.493	ug/L	99
49) Dibromochloromethane	8.808	129	18839	1.210	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102823\
Data File : VU055974.D
Acq On : 28 Oct 2023 15:16
Operator : MD/SY
Sample : 05114-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
C0B58

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/30/2023
Supervised By : Mahesh Dadoda 11/01/2023

Quant Time: Oct 30 04:54:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 30 04:51:15 2023
Response via : Initial Calibration





#13

Acetone

Concen: 5.058 ug/L

RT: 2.641 min Scan# 41

Delta R.T. 0.003 min

Lab File: VU055974.D

Acq: 28 Oct 2023 15:16

Instrument :

MSVOA_U

ClientSampleId :

C0B58

Tgt Ion: 43 Resp: 1404

Ion Ratio Lower Upper

43 100





58 33.5 0.0 69.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/30/2023

Supervised By :Mahesh Dadoda 11/01/2023



#21

2-Butanone

Concen: 2.921 ug/L m

RT: 4.721 min Scan# 1067

Delta R.T. 0.013 min

Lab File: VU055974.D




Acq: 28 Oct 2023 15:16


Tgt Ion: 43 Resp: 11927

Ion Ratio Lower Upper

43 100

72 13.9 14.4 43.4#





#25

Chloroform

Concen: 18.365 ug/L

RT: 5.084 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VU055974.D

Acq: 28 Oct 2023 15:16

Instrument :

MSVOA_U

ClientSampleId :

C0B58

Tgt Ion: 83 Resp: 66356

Ion Ratio Lower Upper

83 100





85 64.2 45.6 84.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/30/2023

Supervised By :Mahesh Dadoda 11/01/2023



#38

Bromodichloromethane

Concen: 4.493 ug/L

RT: 7.103 min Scan# 1808

Delta R.T. 0.000 min

Lab File: VU055974.D

Acq: 28 Oct 2023 15:16




Tgt Ion: 83 Resp: 113786


Ion Ratio Lower Upper

83 100

85 63.7 44.9 83.3

127 9.2 6.6 9.8





#49
 Dibromochloromethane
 Concen: 1.210 ug/L
 RT: 8.808 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VU055974.D
 Acq: 28 Oct 2023 15:16

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B58

Tgt Ion:129 Resp: 18839
 Ion Ratio Lower Upper
 129 100
 127 76.9 53.5 99.5

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/30/2023
 Supervised By :Mahesh Dadoda 11/01/2023

