

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102918\
 Data File : VU027852.D
 Acq On : 29 Oct 2018 13:44
 Operator : MD/SY
 Sample : J5705-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A40C1

Quant Time: Oct 30 03:01:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 30 02:27:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	111267	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	102423	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48540	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	32469	37.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.00%
7) Chloroethane-d5	1.68	69	26377	39.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.14%
11) 1,1-Dichloroethene-d2	2.27	63	47568	30.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.02%
21) 2-Butanone-d5	4.18	46	61955	83.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.98%
24) Chloroform-d	4.65	84	61960	39.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.62%
26) 1,2-Dichloroethane-d4	5.31	65	45342	42.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.14%
32) Benzene-d6	5.34	84	125865	43.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.30%
36) 1,2-Dichloropropane-d6	6.33	67	44765	42.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.98%
41) Toluene-d8	7.57	98	114645	42.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.22%
43) trans-1,3-Dichloropropene-	7.85	79	20168	39.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.76%
47) 2-Hexanone-d5	8.31	63	45331	85.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.06%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	56791	42.23	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.46%
64) 1,2-Dichlorobenzene-d4	11.86	152	42074	42.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU102918\
Data File : VU027852.D
Acq On : 29 Oct 2018 13:44
Operator : MD/SY
Sample : J5705-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A40C1

Quant Time: Oct 30 03:01:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM102218WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 30 02:27:31 2018
Response via : Initial Calibration

