

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102919\
 Data File : VU035542.D
 Acq On : 29 Oct 2019 16:38
 Operator : JC/SP
 Sample : VU1029WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1029WBL01

Quant Time: Oct 30 06:12:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U102919DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Oct 30 05:38:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.16	96	68429	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.67	95	21451	0.87	ug/l	0.00
Spiked Amount 1.000			Recovery	=	87.00%	
68) 1,2-Dichlorobenzene-d4	12.22	152	21200	0.81	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_U\DATA\VU102919\
Data File : VU035542.D
Acq On : 29 Oct 2019 16:38
Operator : JC/SP
Sample : VU1029WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1029WBL01

Quant Time: Oct 30 06:12:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U102919DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Oct 30 05:38:55 2019
Response via : Initial Calibration

