

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102920\
 Data File : VU041010.D
 Acq On : 29 Oct 2020 11:08
 Operator : SY/MD
 Sample : VU1029WBL06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK11

Quant Time: Oct 29 12:43:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 29 12:31:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	113091	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	117998	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.81	152	49504	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	39633	5.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	1.92	69	31249	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.58	65	18260	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.20%
20) 2-Butanone-d5	4.66	46	171345	54.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.08%
24) Chloroform-d	5.08	84	76238	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.72	65	49777	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.74	84	136688	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.70	67	48471	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.91	98	109119	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	8.19	79	19842	4.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.00%
46) 2-Hexanone-d5	8.64	63	116543	50.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.58%
56) 1,1,2,2-Tetrachloroethane-	10.75	84	43436	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.20%
66) 1,2-Dichlorobenzene-d4	12.19	152	43180	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102920\
Data File : VU041010.D
Acq On : 29 Oct 2020 11:08
Operator : SY/MD
Sample : VU1029WBL06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK11

Quant Time: Oct 29 12:43:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 29 12:31:13 2020
Response via : Initial Calibration

