

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU102920\
 Data File : VU041011.D
 Acq On : 29 Oct 2020 11:39
 Operator : SY/MD
 Sample : VU1029WBL07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK12

Quant Time: Oct 29 13:45:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 29 12:31:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	118907	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	120731	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.81	152	51599	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	36257	4.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	1.92	69	29800	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.58	65	18039	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.20%
20) 2-Butanone-d5	4.66	46	168906	50.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.34%
24) Chloroform-d	5.08	84	74369	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.72	65	47866	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	5.74	84	131955	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
36) 1,2-Dichloropropane-d6	6.70	67	46414	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.20%
41) Toluene-d8	7.90	98	104977	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
43) trans-1,3-Dichloropropene-	8.18	79	19766	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	8.64	63	113153	48.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.40%
56) 1,1,2,2-Tetrachloroethane-	10.75	84	43149	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
66) 1,2-Dichlorobenzene-d4	12.19	152	41108	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU102920\
Data File : VU041011.D
Acq On : 29 Oct 2020 11:39
Operator : SY/MD
Sample : VU1029WBL07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK12

Quant Time: Oct 29 13:45:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 29 12:31:13 2020
Response via : Initial Calibration

