


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU102920\
 Data File : VU041008.D
 Acq On : 29 Oct 2020 09:46
 Operator : SY/MD
 Sample : BFB30
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB30

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR102820WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Thu Oct 29 14:46:50 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	14328	PASS
75	95	30	80	57.7	36968	PASS
95	95	100	100	100.0	64085	PASS
96	95	5	9	6.8	4340	PASS
173	174	0.00	2	1.4	652	PASS
174	95	50	120	73.3	47000	PASS
175	174	5	9	7.2	3362	PASS
176	174	95	101	97.1	45616	PASS
177	176	5	9	7.0	3197	PASS

m/z	Abundance
36.00	852.0
37.00	3842.0
38.10	3457.0
39.00	1445.0
40.10	247.0
44.00	877.0
45.10	791.0
46.00	277.0
47.00	927.0
48.00	414.0
49.10	3134.0
50.10	14705.0
51.10	3995.0
52.10	286.0
53.00	111.0
55.10	199.0
56.00	1318.0
57.00	2410.0
60.10	712.0
61.10	3704.0
62.10	4075.0
63.10	3204.0
64.00	532.0
65.10	700.0
66.90	236.0
68.10	7815.0
69.10	7241.0
70.00	631.0
72.00	577.0
73.00	3220.0
74.10	12843.0
75.10	37800.0
76.10	3636.0
77.00	347.0
78.00	342.0
78.90	3189.0
80.00	930.0
81.00	3479.0
81.90	861.0
87.00	2236.0
88.00	2208.0
90.90	274.0
92.10	2151.0
93.10	3348.0
94.10	8283.0
95.10	65720.0
96.10	4619.0
103.90	343.0
105.90	359.0
116.00	300.0
117.00	644.0
118.00	272.0
119.00	593.0
127.90	331.0
129.90	293.0
134.90	152.0
141.00	810.0
143.10	1015.0
143.90	101.0
145.10	347.0
149.80	128.0
154.90	116.0
171.10	210.0
172.10	855.0
173.20	1074.0
174.10	46432.0
175.10	3476.0
176.10	44816.0
177.10	3278.0
178.00	120.0

Instrument :
MSVOA_U
ClientSampleId :
BFB30

m/z	Abundance
36.00	805.0
37.00	3781.0
38.00	3241.0
39.00	1478.0
40.10	239.0
44.00	846.0
45.10	786.0
46.00	201.0
47.00	974.0
48.00	427.0
49.00	3170.0
50.10	14930.0
51.10	4160.0
52.10	263.0
55.10	166.0
56.00	1113.0
57.10	2408.0
60.10	723.0
61.10	3811.0
62.10	3868.0
63.10	3403.0
64.00	530.0
65.10	687.0
67.00	198.0
68.00	7832.0
69.10	7153.0
70.00	565.0
72.00	575.0
73.00	3256.0
74.10	12594.0
75.10	38200.0
76.10	3516.0
77.00	372.0
78.10	331.0
78.90	3169.0
80.00	878.0
81.00	3737.0
81.90	952.0
87.00	2298.0
87.90	2209.0
90.90	279.0
92.10	2075.0
93.10	3209.0
94.10	8093.0
95.10	66088.0
96.10	4467.0
103.90	262.0
105.90	361.0
116.00	340.0
117.00	611.0
118.00	254.0
119.00	539.0
127.90	374.0
128.90	105.0
129.90	237.0
135.00	187.0
141.00	899.0
143.10	1005.0
145.00	320.0
149.80	109.0
154.90	118.0
171.10	172.0
172.10	836.0
173.20	883.0
174.10	48504.0
175.10	3456.0
176.10	47088.0
177.10	3375.0
178.00	139.0

Instrument :
MSVOA_U
ClientSampleId :
BFB30

m/z	Abundance
36.00	688.0
37.00	3513.0
38.00	2792.0
39.00	1319.0
40.10	172.0
44.00	741.0
45.10	622.0
46.00	132.0
47.00	881.0
48.00	381.0
49.00	3045.0
50.10	13349.0
51.10	4011.0
52.10	241.0
53.00	107.0
55.00	162.0
56.00	1010.0
57.10	2007.0
60.00	654.0
61.10	3593.0
62.00	3535.0
63.10	3289.0
64.00	474.0
65.10	644.0
67.00	187.0
68.00	7159.0
69.10	6545.0
70.00	453.0
72.00	512.0
73.00	2989.0
74.10	11317.0
75.10	34904.0
76.10	3149.0
77.00	345.0
78.10	320.0
78.90	2899.0
80.00	764.0
81.00	3468.0
81.90	872.0
87.00	2222.0
88.00	2108.0
90.90	276.0
92.10	1894.0
93.10	2766.0
94.10	7470.0
95.10	60448.0
96.10	3935.0
103.90	251.0
105.90	342.0
116.00	264.0
117.00	511.0
118.00	242.0
119.00	436.0
127.90	344.0
128.90	110.0
130.00	173.0
135.10	214.0
141.00	871.0
143.00	924.0
145.00	333.0
154.90	108.0
161.10	108.0
171.20	102.0
172.00	759.0
174.10	46064.0
175.10	3155.0
176.10	44944.0
177.10	2938.0
178.00	136.0

Instrument :
MSVOA_U
ClientSampleId :
BFB30

Scan 2875 (10.584 min): VU041008.D
BFB30

m/z	Abundance
39.80	113.0
44.00	336.0

Instrument :
MSVOA_U
ClientSampleId :
BFB30