

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045455.D
 Acq On : 29 Oct 2021 11:41
 Operator : SY/MD
 Sample : VU1029WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1029WBL01

Quant Time: Oct 30 04:03:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	183912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	353790	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	340067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	159705	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	154794	51.699	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.400%
35) Dibromofluoromethane	5.292	113	114063	50.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.400%
50) Toluene-d8	7.903	98	450427	51.061	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.120%
62) 4-Bromofluorobenzene	10.636	95	168960	50.264	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045455.D
Acq On : 29 Oct 2021 11:41
Operator : SY/MD
Sample : VU1029WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1029WBL01

Quant Time: Oct 30 04:03:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

