

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045461.D
 Acq On : 29 Oct 2021 14:12
 Operator : SY/MD
 Sample : M4360-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-EB-20211022

Quant Time: Oct 30 04:05:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	192137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	379826	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	371000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	180183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	164772	52.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.360%
35) Dibromofluoromethane	5.295	113	120428	49.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	7.903	98	485742	51.290	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.580%
62) 4-Bromofluorobenzene	10.636	95	184033	50.995	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.980%
Target Compounds						
16) Acetone	2.642	43	2559	1.802	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045461.D
Acq On : 29 Oct 2021 14:12
Operator : SY/MD
Sample : M4360-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-EB-20211022

Quant Time: Oct 30 04:05:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

