

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045468.D
 Acq On : 29 Oct 2021 16:54
 Operator : SY/MD
 Sample : M4406-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CHRT-20310

Quant Time: Oct 30 04:08:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	204319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	398615	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	394985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	197021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	170625	51.295	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.580%
35) Dibromofluoromethane	5.295	113	123992	48.920	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	7.903	98	516965	52.013	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.020%
62) 4-Bromofluorobenzene	10.636	95	200298	52.886	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.780%
Target Compounds						
					Qvalue	
16) Acetone	2.646	43	57913	38.350	ug/l	99
20) Methylene Chloride	3.051	84	8791	2.921	ug/l	97
25) 2-Butanone	4.719	43	13374	5.866	ug/l	99
43) Isopropyl Acetate	5.916	43	20520	2.963	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045468.D
Acq On : 29 Oct 2021 16:54
Operator : SY/MD
Sample : M4406-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CHRT-20310

Quant Time: Oct 30 04:08:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

