

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045469.D
 Acq On : 29 Oct 2021 17:17
 Operator : SY/MD
 Sample : M4395-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MH-FFS

Quant Time: Oct 30 04:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	202332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	404263	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	392283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	194636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	170790	51.848	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.295	113	125444	48.802	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	7.902	98	517826	51.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.740%
62) 4-Bromofluorobenzene	10.636	95	196500	51.158	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.320%
Target Compounds						
						Qvalue
16) Acetone	2.639	43	40214	26.891	ug/l	99
20) Methylene Chloride	3.051	84	8775	2.945	ug/l	98
25) 2-Butanone	4.716	43	11697	5.181	ug/l	95
43) Isopropyl Acetate	5.912	43	29463	4.195	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045469.D
Acq On : 29 Oct 2021 17:17
Operator : SY/MD
Sample : M4395-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MH-FFS

Quant Time: Oct 30 04:08:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

