

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045474.D
 Acq On : 29 Oct 2021 19:13
 Operator : SY/MD
 Sample : M4375-01
 Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-6(1.0-1.5)

Quant Time: Oct 30 04:10:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	218963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	433803	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	423745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	200002	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	180045	50.507	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.020%
35) Dibromofluoromethane	5.292	113	131247	47.582	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.160%
50) Toluene-d8	7.899	98	552250	51.056	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.120%
62) 4-Bromofluorobenzene	10.642	95	210906	51.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045474.D
Acq On : 29 Oct 2021 19:13
Operator : SY/MD
Sample : M4375-01
Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-6(1.0-1.5)

Quant Time: Oct 30 04:10:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

