

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045477.D
 Acq On : 29 Oct 2021 20:22
 Operator : SY/MD
 Sample : M4341-01ME
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WASTEME

Quant Time: Oct 30 04:11:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	220730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	439802	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	432567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	204821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	184964	51.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.940%	
35) Dibromofluoromethane	5.292	113	131751	47.113	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.220%	
50) Toluene-d8	7.899	98	568398	51.833	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.660%	
62) 4-Bromofluorobenzene	10.642	95	222206	53.176	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.360%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.662	96	11309	3.616	ug/l	92
39) Methylcyclohexane	6.761	83	7220	1.387	ug/l	99
68) m/p-Xylenes	9.700	106	28253	4.982	ug/l	98
69) o-Xylene	10.108	106	18015	3.258	ug/l	95
78) n-propylbenzene	10.912	91	20073	1.199	ug/l	100
80) 1,3,5-Trimethylbenzene	11.095	105	58573	5.040	ug/l	99
84) 1,2,4-Trimethylbenzene	11.475	105	146504	12.582	ug/l	100
85) sec-Butylbenzene	11.652	105	23658	1.607	ug/l	98
86) p-Isopropyltoluene	11.799	119	22975	1.943	ug/l	98
89) n-Butylbenzene	12.201	91	56268	5.264	ug/l #	30

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
Data File : VU045477.D
Acq On : 29 Oct 2021 20:22
Operator : SY/MD
Sample : M4341-01ME
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WASTEME

Quant Time: Oct 30 04:11:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

