


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045445.D
 Acq On : 28 Oct 2021 18:18
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Title : SW846 8260
 Last Update : Fri Oct 29 11:09:42 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2877

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	11259	PASS
75	95	30	60	51.9	29645	PASS
95	95	100	100	100.0	57101	PASS
96	95	5	9	7.1	4034	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.9	41637	PASS
175	174	5	9	7.3	3047	PASS
176	174	95	101	97.6	40653	PASS
177	176	5	9	6.3	2577	PASS