


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU102921\
 Data File : VU045453.D
 Acq On : 29 Oct 2021 09:21
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Title : SW846 8260
 Last Update : Fri Oct 29 11:09:42 2021



AutoFind: Scans 2891, 2892, 2893; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	10580	PASS
75	95	30	60	51.0	27267	PASS
95	95	100	100	100.0	53437	PASS
96	95	5	9	6.8	3628	PASS
173	174	0.00	2	1.1	436	PASS
174	95	50	100	71.4	38173	PASS
175	174	5	9	7.4	2836	PASS
176	174	95	101	97.1	37083	PASS
177	176	5	9	6.3	2349	PASS