

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU103018\
 Data File : VU027871.D
 Acq On : 30 Oct 2018 13:26
 Operator : MD/SY
 Sample : VU1030WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK54

Quant Time: Oct 31 02:29:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR103018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 31 02:11:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	97807	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	91137	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	42640	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	32227	4.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.40%
7) Chloroethane-d5	1.68	69	24297	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.40%
11) 1,1-Dichloroethene-d2	2.28	63	45347	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	4.19	46	96123	42.17	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.34%
24) Chloroform-d	4.65	84	54553	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
26) 1,2-Dichloroethane-d4	5.31	65	31674	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.20%
32) Benzene-d6	5.34	84	106018	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
36) 1,2-Dichloropropane-d6	6.33	67	36653	4.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.00%
41) Toluene-d8	7.57	98	99775	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	7.85	79	14354	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.31	63	79420	42.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.92%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	25983	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
64) 1,2-Dichlorobenzene-d4	11.86	152	35303	4.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU103018\
Data File : VU027871.D
Acq On : 30 Oct 2018 13:26
Operator : MD/SY
Sample : VU1030WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK54

Quant Time: Oct 31 02:29:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR103018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 31 02:11:32 2018
Response via : Initial Calibration

