

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U103020\
 Data File : VU041021.D
 Acq On : 29 Oct 2020 18:17
 Operator : SY/MD
 Sample : L4558-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BP6

Quant Time: Oct 30 07:32:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 30 08:06:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	53729	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	54802	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	22084	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	15810	3.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.00%
7) Chloroethane-d5	1.92	69	12907	3.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.40%
11) 1,1-Dichloroethene-d2	2.58	63	24524	2.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.40%#
20) 2-Butanone-d5	4.66	46	81636	54.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.84%
24) Chloroform-d	5.08	84	35239	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.72	65	22943	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	5.74	84	59901	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.80%
36) 1,2-Dichloropropane-d6	6.70	67	21749	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.90	98	47186	3.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.00%
43) trans-1,3-Dichloropropene-	8.19	79	8452	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	8.64	63	53754	48.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.54%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	20307	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	19370	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	646	0.126 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU103020\
Data File : VU041021.D
Acq On : 29 Oct 2020 18:17
Operator : SY/MD
Sample : L4558-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BP6

Quant Time: Oct 30 07:32:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 30 08:06:28 2020
Response via : Initial Calibration

