

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U103020\
 Data File : VU041019.D
 Acq On : 29 Oct 2020 17:30
 Operator : SY/MD
 Sample : L4558-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BQ4

Quant Time: Oct 30 07:24:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 30 08:06:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	53291	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	55059	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	22445	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	16299	3.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.80%
7) Chloroethane-d5	1.92	69	13156	3.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.40%
11) 1,1-Dichloroethene-d2	2.58	63	24543	2.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.80%#
20) 2-Butanone-d5	4.66	46	81908	55.56	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.12%
24) Chloroform-d	5.08	84	34949	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.72	65	23479	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.74	84	62192	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
36) 1,2-Dichloropropane-d6	6.70	67	22182	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.91	98	47400	3.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.00%
43) trans-1,3-Dichloropropene-	8.19	79	8834	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.64	63	54832	49.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.02%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	21267	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	19439	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU103020\
Data File : VU041019.D
Acq On : 29 Oct 2020 17:30
Operator : SY/MD
Sample : L4558-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BQ4

Quant Time: Oct 30 07:24:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 30 08:06:28 2020
Response via : Initial Calibration

