

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U103020\
 Data File : VU041025.D
 Acq On : 29 Oct 2020 19:50
 Operator : SY/MD
 Sample : L4558-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BQ0

Quant Time: Oct 30 07:53:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 30 08:06:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	49483	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	51109	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	20708	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	15290	3.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.60%
7) Chloroethane-d5	1.92	69	12999	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.40%
11) 1,1-Dichloroethene-d2	2.58	63	23435	2.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.40%#
20) 2-Butanone-d5	4.66	46	79532	58.10	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.20%
24) Chloroform-d	5.08	84	34383	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.72	65	22121	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
32) Benzene-d6	5.74	84	58229	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
36) 1,2-Dichloropropane-d6	6.70	67	21777	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.91	98	44501	3.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.80%
43) trans-1,3-Dichloropropene-	8.19	79	8132	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.64	63	53751	51.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.50%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	20019	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	18064	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU103020\
Data File : VU041025.D
Acq On : 29 Oct 2020 19:50
Operator : SY/MD
Sample : L4558-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BQ0

Quant Time: Oct 30 07:53:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 30 08:06:28 2020
Response via : Initial Calibration

