

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103023\
 Data File : VU055992.D
 Acq On : 30 Oct 2023 13:11
 Operator : MD/SY
 Sample : 05113-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B43

Quant Time: Oct 31 05:30:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 31 05:28:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	325769	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	319811	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	150752	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	94060	3.650	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.000%
7) Chloroethane-d5	1.914	69	86074	4.285	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.800%
11) 1,1-Dichloroethene-d2	2.566	65	43828	4.006	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.200%
20) 2-Butanone-d5	4.627	46	227411	49.466	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.940%
24) Chloroform-d	5.062	84	215289	4.508	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.698	65	102351	4.399	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
32) Benzene-d6	5.724	84	436192	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.689	67	135850	4.810	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.894	98	365813	4.233	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.180	79	44345	4.015	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.631	63	183106	65.306	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	130.620%#
56) 1,1,2,2-Tetrachloroeth...	10.753	84	100812	4.902	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	135910	4.832	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103023\
Data File : VU055992.D
Acq On : 30 Oct 2023 13:11
Operator : MD/SY
Sample : 05113-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B43

Quant Time: Oct 31 05:30:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 31 05:28:08 2023
Response via : Initial Calibration

