

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103122\
 Data File : VU051611.D
 Acq On : 31 Oct 2022 15:04
 Operator : JC/MD
 Sample : N5319-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAA9

Quant Time: Nov 01 01:27:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 01 01:25:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	177137	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	187835	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	72002	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	62722	4.701	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.000%	
7) Chloroethane-d5	1.909	69	74936	5.386	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.800%	
11) 1,1-Dichloroethene-d2	2.565	65	23026	4.756	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	95.200%	
20) 2-Butanone-d5	4.616	46	279201	57.787	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	115.580%	
24) Chloroform-d	5.060	84	168559	5.399	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	108.000%	
26) 1,2-Dichloroethane-d4	5.700	65	95286	5.650	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	113.000%	
32) Benzene-d6	5.726	84	296133	4.996	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.000%	
36) 1,2-Dichloropropane-d6	6.687	67	111449	5.377	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.600%	
41) Toluene-d8	7.896	98	197776	4.243	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.800%	
43) trans-1,3-Dichloroprop...	8.179	79	24094	3.802	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.000%	
46) 2-Hexanone-d5	8.632	63	189540	49.681	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.360%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	105297	5.544	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.800%	
66) 1,2-Dichlorobenzene-d4	12.192	152	74774	5.658	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU103122\
Data File : VU051611.D
Acq On : 31 Oct 2022 15:04
Operator : JC/MD
Sample : N5319-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAA9

Quant Time: Nov 01 01:27:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 01 01:25:31 2022
Response via : Initial Calibration

