

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110218\
 Data File : VU027922.D
 Acq On : 02 Nov 2018 13:40
 Operator : MD/SY
 Sample : J5192-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 AU-05-103118-A

Quant Time: Nov 03 05:02:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190028	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	334149	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	295538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	123847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	145314	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
35) Dibromofluoromethane	4.89	113	108062	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
50) Toluene-d8	7.57	98	424120	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
62) 4-Bromofluorobenzene	10.31	95	140711	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
Target Compounds						
16) Acetone	2.32	43	12331	7.035	ug/l	99
18) Methyl Acetate	2.62	43	4822	1.000	ug/l	92
20) Methylene Chloride	2.70	84	19306	7.543	ug/l	95
43) Isopropyl Acetate	5.55	43	185800	23.128	ug/l	99
95) Naphthalene	13.75	128	71476	6.879	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110218\
Data File : VU027922.D
Acq On : 02 Nov 2018 13:40
Operator : MD/SY
Sample : J5192-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
AU-05-103118-A

Quant Time: Nov 03 05:02:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

