

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110218\
 Data File : VU027937.D
 Acq On : 02 Nov 2018 19:33
 Operator : MD/SY
 Sample : J5767-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CL-02-110118-D

Quant Time: Nov 03 06:03:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	335238	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	297373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	125063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	147279	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
35) Dibromofluoromethane	4.88	113	110323	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
50) Toluene-d8	7.57	98	429969	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	10.31	95	139034	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
Target Compounds						
16) Acetone	2.32	43	7815	4.482	ug/l	100
18) Methyl Acetate	2.62	43	5021	1.047	ug/l #	78
20) Methylene Chloride	2.70	84	7957	3.125	ug/l	86
43) Isopropyl Acetate	5.55	43	191123	23.713	ug/l	100
95) Naphthalene	13.75	128	21132	1.885	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110218\
Data File : VU027937.D
Acq On : 02 Nov 2018 19:33
Operator : MD/SY
Sample : J5767-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CL-02-110118-D

Quant Time: Nov 03 06:03:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

