


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110220\
 Data File : VU041041.D
 Acq On : 02 Nov 2020 10:23
 Operator : SY/MD
 Sample : BFB03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB03

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Nov 03 03:57:56 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	17346	PASS
75	95	30	80	57.0	45040	PASS
95	95	100	100	100.0	79032	PASS
96	95	5	9	6.3	5015	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	72.5	57314	PASS
175	174	5	9	7.8	4459	PASS
176	174	95	101	95.5	54744	PASS
177	176	5	9	6.3	3450	PASS

m/z	Abundance
44.00	159.0

Instrument :
MSVOA_U
ClientSampleId :
BFB03

m/z	Abundance
36.10	897.0
37.00	4988.0
38.10	4287.0
39.10	1838.0
44.00	719.0
45.00	839.0
46.00	117.0
47.10	978.0
48.00	575.0
49.10	3748.0
50.10	17152.0
51.10	5460.0
52.10	205.0
55.00	267.0
56.10	1529.0
57.10	2541.0
60.00	862.0
61.10	4426.0
62.10	4634.0
63.00	3585.0
64.10	568.0
65.00	505.0
67.00	182.0
68.10	8830.0
69.10	9139.0
70.00	651.0
72.00	539.0
73.10	4098.0
74.10	15275.0
75.10	44264.0
76.10	3853.0
77.10	495.0
78.00	333.0
79.00	3988.0
80.00	963.0
81.00	3896.0
81.90	947.0
83.00	100.0
87.10	3049.0
88.00	2923.0
91.00	345.0
92.10	2693.0
93.10	3617.0
94.10	9649.0
95.10	77504.0
96.10	4955.0
97.00	216.0
104.00	530.0
104.90	105.0
105.90	385.0
115.00	132.0
116.00	365.0
117.00	679.0
117.90	295.0
119.00	460.0
128.00	341.0
129.00	114.0
130.00	248.0
131.00	110.0
135.00	168.0
141.00	838.0
142.10	100.0
143.00	1264.0
145.00	324.0
146.00	114.0
147.90	196.0
149.00	104.0
155.00	155.0
157.00	124.0
171.10	157.0
172.10	785.0
174.10	54264.0
175.10	4239.0
176.10	52000.0
177.10	3218.0
178.00	108.0

Instrument :
MSVOA_U
ClientSampleId :
BFB03

m/z	Abundance
36.10	903.0
37.00	5178.0
38.10	4395.0
39.10	1943.0
44.00	780.0
45.00	840.0
46.20	113.0
47.10	1036.0
48.00	579.0
49.00	3884.0
50.10	17904.0
51.10	5697.0
52.10	225.0
55.00	259.0
56.10	1669.0
57.10	2569.0
60.00	920.0
61.10	4528.0
62.10	4824.0
63.10	3683.0
64.10	509.0
65.00	501.0
67.00	192.0
68.10	9288.0
69.10	9470.0
70.10	636.0
72.00	578.0
73.10	4147.0
74.10	15948.0
75.10	46568.0
76.10	3930.0
77.00	536.0
78.00	349.0
79.00	3962.0
80.00	1077.0
81.00	3993.0
82.00	965.0
87.10	3260.0
88.10	3233.0
91.00	382.0
92.10	2745.0
93.10	3852.0
94.10	10225.0
95.10	81704.0
96.10	5137.0
97.00	200.0
104.00	537.0
105.00	110.0
105.90	416.0
111.00	118.0
115.00	113.0
116.00	400.0
117.00	739.0
117.90	325.0
119.00	527.0
128.00	299.0
129.00	116.0
130.00	272.0
131.00	115.0
135.10	181.0
136.90	121.0
141.00	949.0
142.10	104.0
143.00	1298.0
145.00	298.0
146.00	113.0
147.90	188.0
155.00	159.0
157.00	145.0
171.10	143.0
172.10	766.0
174.10	59288.0
175.10	4611.0
176.10	56648.0
177.10	3538.0
178.00	107.0

Instrument :
MSVOA_U
ClientSampleId :
BFB03

m/z	Abundance
36.10	790.0
37.00	4734.0
38.00	4182.0
39.10	1877.0
44.00	746.0
45.10	758.0
46.30	104.0
47.10	939.0
48.00	519.0
49.00	3753.0
50.10	16984.0
51.10	5347.0
52.10	210.0
55.00	246.0
56.10	1612.0
57.10	2431.0
58.00	109.0
60.00	928.0
61.10	4351.0
62.10	4519.0
63.10	3469.0
64.10	413.0
65.10	477.0
67.00	185.0
68.10	8723.0
69.10	8858.0
70.10	612.0
72.00	543.0
73.10	3912.0
74.10	15057.0
75.10	44288.0
76.10	3597.0
77.00	490.0
77.90	400.0
79.00	3665.0
80.00	1051.0
81.00	3695.0
82.00	914.0
87.00	3176.0
88.10	3177.0
90.90	401.0
92.10	2541.0
93.10	3754.0
94.10	9940.0
95.10	77888.0
96.10	4953.0
97.00	130.0
103.90	461.0
105.00	105.0
105.90	422.0
111.00	137.0
116.00	365.0
117.00	728.0
117.90	336.0
119.00	548.0
128.00	230.0
129.00	112.0
130.00	261.0
131.00	102.0
135.00	176.0
136.80	146.0
141.00	957.0
142.00	106.0
143.00	1173.0
145.10	251.0
146.00	113.0
148.00	177.0
155.00	156.0
157.00	149.0
159.00	112.0
171.00	110.0
172.00	654.0
174.10	58392.0
175.10	4527.0
176.10	55584.0
177.10	3594.0

Instrument :
MSVOA_U
ClientSampleId :
BFB03